

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod Lisem 129, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1160643

Product: Active flexible sensor for measurement of alternating current

Type: AMOS M

Rating: 230 V AC, current ranges 10; 30; 100; 300 or 1000A, safety class II, overvoltage category IV,
supply voltage ± 5 - 12 V DC, output voltage 1,0 V AC at Inom

Ordering firm: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Manufacturer: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Factories: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 602805-01/01 of: 15.08.2016, 302970-01/02 of: 11.07.2013

A sample of the product was found to be in conformity with:
EN 61010-1:10, EN 61010-2-032:12, EN 61326-1:13

Other data:

The validity of the certificate is limited to: 31.08.2019

31.08.2016

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



Stamp



602805-01