

# ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC  
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHJECHISCHE REPUBLIK  
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÈQUE  
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod Lisem 129, 171 02 Praha 8 - Troja

## CERTIFICATE

**No.:** 1170834

**Product:** Active flexible sensor for measurement of alternating current

**Type:** AMOS M long

**Rating:** 30, 100, 300, 1000, 3000 A, Safety class II, CAT IV/300V or CAT III/600V  
Supply voltage  $\pm 3$  V to  $\pm 12$  V DC, Output voltage 1,0 V AC for Inom

**Ordering firm:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Manufacturer:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Factories:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Trade mark:** MEgA

**The test results are stated in the test-report No.:** 704202-01/01 of: 22.11.2017

**A sample of the product was found to be in conformity with:**  
IEC 61010-1:2010+A1:2016  
IEC 61010-2-032:2012

**Other data:**

**The validity of the certificate is limited to:** 30.11.2020

23.11.2017

Prague

RNDr. Milan Press  
Deputy Head of Certification Body



Stamp



704202-01